Cleaning and Disinfection Guidelines

Guidance on hygienic use

By choosing to use our products, you have taken an important step towards ensuring hygienic conditions in your working environment. To make it easier for you to work with these devices, to maintain their functionality in the long term and to comply with hygiene requirements, we would like to offer you some guidance below.

The hygienic silicone mouse mat from Baaske Medical is characterised by its ultra-thin, non-slip design and its pleasant feel. The finely textured surface allows for precise control, even with high-resolution laser mice, and is particularly easy to clean and disinfect.

The method and frequency of cleaning and disinfection required depend on the specific area of use and the particular hygiene requirements of the facility. Your hygiene officer can provide you with tailored advice on drawing up local hygiene plans. The silicone mouse mat can be cleaned with water, soap, standard disinfectants or 10 per cent bleach. It can also be processed in a thermal disinfector or autoclave at temperatures of up to 108 °C. Always follow the manufacturers' instructions for use and the applicable hygiene guidelines of your facility.

Cleaning and Disinfection Guidelines

Cleaning / Disinfection:

For routine cleaning, we recommend wiping the product with a soft microfibre cloth dampened with lukewarm water and a small amount of a mild, pH-neutral cleaning agent. Then wipe the product with a clean, damp cloth and leave it to air-dry flat at room temperature.

- The product can be disinfected. Use suitable surface disinfectants for this purpose. Before using the selected disinfectant for the first time, test it on an inconspicuous corner of the mousepad to check material compatibility.
- Alcohol-based disinfectants, including 70% isopropyl alcohol (IPA), should only be used after a prior compatibility test, as these may affect the colour, gloss, adhesion or surface finish of the product.
- The silicone mouse pad is suitable for thermal reprocessing at temperatures of up to 108 °C in an autoclave or thermal disinfector.
- Apply cleaning and disinfectant solutions to a soft cloth and wipe the surface with it. Do not spray any solutions directly onto the product and avoid prolonged soaking.

Avoid abrasive sponges, strongly alkaline or acidic cleaning agents, acetone, thinners and bleach.

For regular use, we recommend gentle cleaning with a mild, pH-neutral cleaning agent.