Cleaning and Disinfection Guidelines

Guidance on hygienic use

By choosing to use our products, you have taken an important step towards ensuring hygienic conditions in your working environment. To make it easier for you to work with this equipment, to maintain its functionality in the long term and to comply with hygiene requirements, we would like to provide you with the following guidance.

Please note the information regarding the protection rating of the type of device you are using. Before you start cleaning, ensure that pressing any buttons or controls during the cleaning process will not trigger any unintended actions. If necessary, temporarily switch to a non-critical application. It is recommended that you carry out cleaning and disinfection whilst the computer is switched off.

The type and frequency of the cleaning and disinfection required depend on the specific area of use and the particular hygiene requirements of the facility. Your hygiene officer can provide you with tailored advice on drawing up local hygiene plans. Due to the wide variety of substances available on the market and constant market fluctuations, we are unfortunately unable to provide a binding recommendation regarding the disinfection of our products. Consequently, we cannot accept any liability for the use and application of these substances. In principle, we can only offer guidance regarding the substances to be used, namely an aldehyde-free, ethanol-based rapid disinfection method.

Cleaning and Disinfection Guidelines

Cleaning / Disinfection:

Wipe the product down with a dry, clean cloth and follow the manufacturer's instructions or hospital protocol.

• The product and its peripherals must not be immersed in water or rinsed. If you accidentally spill liquid on the device, disconnect it from the power supply. Contact your IT adviser regarding the device's continued safety before putting it back into service.

• Do not spray cleaning agents onto the housing.

• Do not use disinfectants containing phenol.

• Do not autoclave the product or its peripherals, nor clean them with strong aromatic, chlorinated, ketone-based, ether-based or ester-based solvents, sharp tools or abrasives. Never immerse electrical connectors in water or other liquids.

We are, of course, happy to answer any questions you may have.

Cleaning and Disinfection Guidelines

Cleaning and disinfectant products	Baaske Medical Panel PC
CIDEX	✓
Control III Disinfectant Germicide	✓
Caviwipes	✓
Dispatch Disinfectant Cleaner CLH69101	✓
Puregreen 24 Disinfectant	✓
Alcohol	✓
Isopropyl (Isopropanol) alcohol	✓
Aceton	✓
LpH se	✓
Bleach	✓
Chloride	✓
Actichlor	✓
Betasept	✓
Mikrozid	✓
Viraguard	✓
Windex	✓
Incidin	✓
Virkon	✓
SDW 70C	✓
Trigene	✓
Antigone	✓
SporeClear	✓
MikroBac	✓
Peridox	✓
3M HB Quat	✓
Sodium hypochlorite	✓